

NEMA MS 1-2008 (R2014, R2020)

Standard for
Determination of
Signal-to-Noise Ratio
(SNR) in Diagnostic
Magnetic Resonance
Imaging



NEMA Standards Publication MS 1-2008 (R2014, R2020)

*Determination of Signal-to-Noise Ratio (SNR)
in Diagnostic Magnetic Resonance Imaging*

Published by:

National Electrical Manufacturers Association

1300 North 17th Street, Suite 900

Rosslyn, VA 22209

www.nema.org

© 2021 National Electrical Manufacturers Association. All rights including translation into other languages, reserved under the Universal Copyright Convention, the Berne Convention for the Protection of Literary and Artistic Works, and the International and Pan American Copyright Conventions.

NOTICE AND DISCLAIMER

The information in this publication was considered technically sound by the consensus of persons engaged in the development and approval of the document at the time it was developed. Consensus does not necessarily mean that there is unanimous agreement among every person participating in the development of this document.

NEMA Standards and guideline publications, of which the document contained herein is one, are developed through a voluntary consensus Standards development process. This process brings together volunteers and/or seeks out the views of persons who have an interest in the topic covered by this publication. While NEMA administers the process and establishes rules to promote fairness in the development of consensus, it does not write the document and it does not independently test, evaluate, or verify the accuracy or completeness of any information or the soundness of any judgments contained in its Standards and guideline publications.

NEMA disclaims liability for any personal injury, property, or other damages of any nature whatsoever, whether special, indirect, consequential, or compensatory, directly or indirectly resulting from the publication, use of, application, or reliance on this document. NEMA disclaims and makes no guaranty or warranty, express or implied, as to the accuracy or completeness of any information published herein, and disclaims and makes no warranty that the information in this document will fulfill any of your particular purposes or needs. NEMA does not undertake to guarantee the performance of any individual manufacturer or seller's products or services by virtue of this Standard or guide.

In publishing and making this document available, NEMA is not undertaking to render professional or other services for or on behalf of any person or entity, nor is NEMA undertaking to perform any duty owed by any person or entity to someone else. Anyone using this document should rely on his or her own independent judgment or, as appropriate, seek the advice of a competent professional in determining the exercise of reasonable care in any given circumstances. Information and other Standards on the topic covered by this publication may be available from other sources, which the user may wish to consult for additional views or information not covered by this publication.

NEMA has no power, nor does it undertake to police or enforce compliance with the contents of this document. NEMA does not certify, test, or inspect products, designs, or installations for safety or health purposes. Any certification or other statement of compliance with any health or safety-related information in this document shall not be attributable to NEMA and is solely the responsibility of the certifier or maker of the statement.

Contents

		Page
	Preamble	ii
	Foreword.....	iii
	Rationale.....	iv
	Scope.....	iv
Section 1	DEFINITIONS	1
1.1	Specification Volume	1
1.2	Specification Area.....	1
1.3	Measurement Region of Interest (MROI)	1
1.4	Image Signal.....	1
1.5	Image Noise.....	1
1.6	Image Signal-to-Noise Ratio.....	1
1.7	Image Artifact.....	1
1.8	Voxel Dimensions	1
1.9	Real Image	2
1.10	Magnitude Image.....	2
Section 2	METHODS OF MEASUREMENT	3
2.1	Test Hardware	3
2.1.1	Size of the Signal Producing Volume.....	3
2.1.2	RF Coil Loading Characteristics.....	3
2.1.3	MR Characteristics of the Signal Producing Volume	3
2.2	Scan Conditions	3
2.3	Measurement Procedure	4
2.3.1	Image Signal Determination.....	4
2.3.2	Procedures to Evaluate Image Noise	4
2.3.3	Procedure to Determine SNR	9
Section 3	RESULTS	10
3.1	Reporting of Results	10
3.1.1	Geometric and Phantom Information	10
3.1.2	Data Acquisition Parameters.....	10
3.1.3	SNR Results.....	11
3.1.4	Reconstruction Parameters.....	11
3.1.5	Additional Data	11
3.1.6	Repeatability Data	11
3.1.7	Sources of Error	11
Appendices		
A	An Alternative Measure of Noise Standard Deviation	12
B	Changes to Standard.....	13
Table		
3-1	Data Acquisition Parameters	10
Figures		
1-1	A Schematic Showing the Relative Location of Defined Terms	2
2-1	A Schematic Showing Example Background Noise Measurement Region Locations for Method 4.....	9

Preamble

This is one of a series of test Standards developed by the medical diagnostic imaging industry for the measurement of performance parameters governing image quality of magnetic resonance (MR) imaging systems. These test Standards are intended for the use of equipment manufacturers, prospective purchasers, and users alike.

Manufacturers are permitted to use these Standards for the determination of system performance specifications. This standardization of performance specifications is of benefit to the prospective equipment purchaser, and the parameters supplied with each NEMA measurement serve as a guide to those factors that can influence the measurement. These Standards can also serve as reference procedures for acceptance testing and periodic quality assurance.

It must be recognized, however, that not all test Standards lend themselves to measurement at the installation site. Some test Standards require instrumentation better suited to factory measurements, while others require the facilities of an instrumentation laboratory to assure stable test conditions necessary for reliable measurements.

The NEMA test procedures are carried out using the normal clinical operating mode of the system. For example, Standard calibration procedures, Standard clinical sequences, and Standard reconstruction processes shall be used. No modifications to alter test results shall be used unless otherwise specified in these Standards.

The NEMA Magnetic Resonance Section has identified a set of key magnetic resonance image quality parameters. This Standards publication describes the measurement of one of these parameters.

Equivalence

It is intended and expected that manufacturers or others who claim compliance with these NEMA Standard test procedures for the determination of image quality parameters shall have carried out the tests in accordance with the procedures specified in the published Standards.

In those cases where it is impossible or impractical to follow the literal prescription of a NEMA test procedure, a complete description of any deviation from the published procedure must be included with any measurement claimed equivalent to the NEMA Standard. The validity or equivalence of the modified procedure will be determined by the reader.

Uncertainty of the Measurements

The measurement uncertainty of the image quality parameter determined using this Standards publication is to be reported, together with the value of the parameter. Justification for the claimed uncertainty limits shall also be provided by a listing and discussion of sources and magnitudes of error.

Foreword

This Standards publication is classified as a NEMA Standard unless otherwise noted. It describes four methods to measure image signal-to-noise ratio (SNR). It is intended for use by MRI system manufacturers, manufacturers of accessory equipment (including radiofrequency coils), and by MRI end users.

The major feature of the first method is that the SNR performance of the system is evaluated using a Standard clinical scan procedure. However, it should be noted that since this method involves the subtraction of two images, it can be very sensitive to system instabilities that may occur during the data acquisition process. If results are highly variable, it is advisable to perform the alternative calculation of Standard deviation, described in method #1, or use methods #2, #3, or #4. These alternative methods have been designed to be less susceptible to system instabilities and can be used to determine if any variability in the SNR is due to system instability or genuinely poor SNR. Both methods are intended to measure thermal and other broadband, non-structured noise, and specifically do not address low frequency variations in an image or artifacts as defined herein.

This Standards publication has been developed by the Magnetic Resonance Section of the National Electrical Manufacturers Association. User needs have been considered throughout the development of this publication. Proposed or recommended revisions should be submitted to:

Vice-President, Engineering Department
National Electrical Manufacturers Association
1300 North 17th Street, Suite 1752
Rosslyn, VA 22209

This Standards publication was developed by the Magnetic Resonance Section. Section approval of the Standard does not necessarily imply that all section Members voted for its approval or participated in its development. At the time it was approved, the Magnetic Resonance Section was composed of the following Members:

Computer Imaging Reference Systems—Norfolk, VA
GE Healthcare, Inc.—Milwaukee, WI
Hitachi Medical Systems America, Inc.—Twinsburg, OH
Invivo—Gainesville, FL
Philips Healthcare—Andover MA
Siemens Medical Solutions, Inc.—Malvern, PA
Toshiba America Medical Systems—Tustin, CA

Rationale

Image SNR is a parameter that relates to clinical usefulness of magnetic resonance images and also is a sensitive measure of hardware performance. Experience has shown that variations in system calibration, gain, coil tuning, radiofrequency shielding, or other similar parameters are usually demonstrated by a corresponding change in image SNR.

Scope

This document defines methods for measuring the signal-to-noise ratio of magnetic resonance images obtained under a specific set of conditions and using single-channel volume receiver coils. This document does not address the use of special purpose coils (see MS 6) or coils that employ multiple receiver channels for operation (see MS 9).