

NEMA Standards Publication

NEMA NU 1-2012

Performance Measurements of Gamma Cameras

National Electrical Manufacturers Association



NEMA Standards Publication NU 1-2012

Performance Measurements of Gamma Cameras

Published by:

National Electrical Manufacturers Association

1300 North 17th Street, Suite 900

Rosslyn, Virginia 22209

www.nema.org

www.medicalimaging.org

© 2013 National Electrical Manufacturers Association. All rights, including translation into other languages, reserved under the Universal Copyright Convention, the Berne Convention for the Protection of Literary and Artistic Works, and the International and Pan American copyright conventions.

NOTICE AND DISCLAIMER

The information in this publication was considered technically sound by a consensus among persons engaged in its development at the time it was approved. Consensus does not necessarily mean there was unanimous agreement among every person participating in the development process.

The National Electrical Manufacturers Association (NEMA) standards and guideline publications, of which the document herein is one, are developed through a voluntary standards development process. This process brings together volunteers and/or seeks out the views of persons who have an interest in the topic covered by this publication. Although NEMA administers the process and establishes rules to promote fairness in the development of consensus, it does not write the documents, nor does it independently test, evaluate, or verify the accuracy or completeness of any information or the soundness of any judgments contained in its standards and guideline publications.

NEMA disclaims liability for any personal injury, property, or other damages of any nature, whether special, indirect, consequential, or compensatory, directly or indirectly resulting from the publication, use of, application, or reliance on this document. NEMA disclaims and makes no guaranty or warranty, express or implied, as to the accuracy or completeness of any information published herein, and disclaims and makes no warranty that the information in this document will fulfill any particular purpose(s) or need(s). NEMA does not undertake to guarantee the performance of any individual manufacturer's or seller's products or services by virtue of this standard or guide.

In publishing and making this document available, NEMA is not undertaking to render professional or other services for or on behalf of any person or entity, nor is NEMA undertaking to perform any duty owed by any person or entity to someone else. Anyone using this document should rely on his or her own independent judgment or, as appropriate, seek the advice of a competent professional in determining the exercise of reasonable care in any given circumstance. Information and other standards on the topic covered by this publication may be available from other sources, which the user may wish to consult for additional views or information not covered by this publication.

NEMA has no power, nor does it undertake to police or enforce compliance with the contents of this document. NEMA does not certify, test, or inspect products, designs, or installations for safety or health purposes. Any certification or other statement of compliance with any health- or safety-related information in this document shall not be attributable to NEMA and is solely the responsibility of the certifier or maker of the statement.

CONTENTS

	Page
SECTION 1 SCOPE.....	1
1.1 DEFINITIONS	1
1.2 REFERENCES.....	4
1.3 TEST EQUIPMENT, CONDITIONS, AND RESULTS.....	4
1.3.1 Source Holders and Test Fixtures	4
1.3.2 Radiation Sources	4
1.3.3 Test Conditions	5
1.3.4 Reporting.....	5
SECTION 2 TESTS OF GAMMA CAMERA DETECTORS.....	6
2.1 INTRINSIC SPATIAL RESOLUTION.....	6
2.1.1 Test Conditions	6
2.1.2 Test Equipment.....	7
2.1.3 Measurement Procedure	7
2.1.4 Calculations and Analysis	8
2.1.5 Reporting.....	10
2.2 INTRINSIC SPATIAL LINEARITY	10
2.2.1 Test Conditions	10
2.2.2 Test Equipment.....	10
2.2.3 Measurement Procedure	10
2.2.4 Calculations and Analysis	11
2.2.5 Reporting.....	11
2.3 INTRINSIC ENERGY RESOLUTION	11
2.3.1 Test Conditions	12
2.3.2 Test Equipment.....	12
2.3.3 Measurement Procedure	12
2.3.4 Calculations and Analysis	12
2.3.5 Reporting.....	13
2.4 INTRINSIC FLOOD FIELD UNIFORMITY	13
2.4.1 Test Conditions	13
2.4.2 Test Equipment.....	13
2.4.3 Measurement Procedure	13
2.4.4 Calculations and Analysis	14
2.4.5 Reporting.....	15
2.5 MULTIPLE WINDOW SPATIAL REGISTRATION	15
2.5.1 Test Conditions	16
2.5.2 Test Equipment.....	16
2.5.3 Measurement Procedure	16
2.5.4 Calculations and Analysis	17
2.5.5 Reporting.....	18
2.6 INTRINSIC COUNT RATE PERFORMANCE IN AIR	18
2.6.1 Decaying Source Method.....	18
2.6.2 Intrinsic Count Rate Performance in Air—Copper Plates Method	21
2.7 INTRINSIC SPATIAL RESOLUTION AT 75,000 COUNTS PER SECOND.....	25
2.8 INTRINSIC FLOOD FIELD UNIFORMITY AT 75,000 COUNTS PER SECOND.....	25
SECTION 3 TESTS OF GAMMA CAMERA DETECTORS WITH S.....	26
3.1 SYSTEM SPATIAL RESOLUTION WITHOUT SCATTER	26
3.1.1 Test Conditions	26
3.1.2 Test Equipment.....	26

3.1.3	Measurement Procedure	26
3.1.4	Calculations and Analysis	27
3.1.5	Reporting	27
3.2	SYSTEM SPATIAL RESOLUTION WITH SCATTER	27
3.2.1	Test Conditions	27
3.2.2	Test Equipment.....	27
3.2.3	Measurement Procedure	27
3.2.4	Calculations and Analysis	28
3.2.5	Reporting	28
3.3	SYSTEM PLANAR SENSITIVITY AND COLLIMATOR PENETRATION AND SCATTER	28
3.3.1	Test Conditions	28
3.3.2	Test Equipment.....	29
3.3.3	Measurement Procedure	29
3.3.4	Calculations and Analysis	29
3.3.5	Reporting	31
3.3.6	Referenced Documents	31
3.4	DETECTOR SHIELDING	31
3.4.1	Test Conditions	31
3.4.2	Test Equipment.....	32
3.4.3	Measurement Procedure	32
3.4.4	Calculations and Analysis	33
3.4.5	Reporting	34
3.5	SYSTEM COUNT RATE PERFORMANCE WITH SCATTER.....	34
3.5.1	Test Conditions	34
3.5.2	Test Equipment.....	34
3.5.3	Measurement Procedure	35
3.5.4	Calculations and Analysis	36
3.5.5	Reporting	36
SECTION 4 TESTS SPECIFIC TO GAMMA CAMERA TOMOGRAPHIC SYSTEMS		37
4.1	SYSTEM ALIGNMENT	37
4.1.1	Test Conditions	37
4.1.2	Test Equipment.....	37
4.1.3	Measurement Procedure	38
4.1.4	Calculations and Analysis	38
4.1.5	Reporting	40
4.2	SPECT RECONSTRUCTED SPATIAL RESOLUTION WITHOUT SCATTER.....	40
4.2.1	Test Conditions	40
4.2.2	Test Equipment.....	40
4.2.3	Measurement Procedure	40
4.2.4	Calculations and Analysis	40
4.2.5	Reporting	42
4.3	SPECT RECONSTRUCTED SPATIAL RESOLUTION WITH SCATTER	42
4.3.1	Test Conditions	42
4.3.2	Test Equipment.....	42
4.3.3	Measurement Procedure	43
4.3.4	Calculations and Analysis	44
4.3.5	Reporting	44
4.4	SYSTEM VOLUME SENSITIVITY	44
4.4.1	Test Conditions	45
4.4.2	Test Equipment.....	45
4.4.3	Measurement Procedure	45
4.4.4	Calculations and Analysis	46
4.4.4	Reporting	46

4.5	DETECTOR-DETECTOR SENSITIVITY VARIATION	46
4.5.1	Test Equipment.....	47
4.5.2	Measurement Procedure	47
4.5.3	Calculations and Analysis	47
4.5.4	Reporting.....	47
SECTION 5 TESTS SPECIFIC TO GAMMA CAMERA WHOLE-BODY SCANNING SYSTEMS.....		48
5.1	WHOLE-BODY SYSTEM SPATIAL RESOLUTION WITHOUT SCATTER	48
5.1.1	Test Conditions	48
5.1.2	Test Equipment.....	48
5.1.3	Measurement Procedure	49
5.1.4	Calculation and Analysis.....	49
5.1.5	Reporting.....	50
5.1.6	Rationale.....	50

FIGURES

	Page
Figure 2–1 Collimated Source Geometry	7
Figure 2–2 Lead Masks for Measurement of Spatial Resolution and Linearity.....	8
Figure 2–3 Determination of FWHM and FWTM	9
Figure 2–4 Cylindrical Source Holder for Multiple Window Spatial Registration Measurement Showing Liquid Ga-67 Source Inside	16
Figure 2–5 Source Holder for Count Rate Measurements.....	19
Figure 2–6 Example Count Rate Curve	21
Figure 3–1 Source Positions for Shield Leakage Measurements	32
Figure 3–2 Phantom for Measuring Count Rate Performance in Scatter	35
Figure 4–1 Positions of the Three Co-Planar Point Sources for Measuring SPECT Head Alignment and Reconstructed System Resolution without Scatter.....	38
Figure 4–2 SPECT Reconstructed Spatial Resolution with Scatter.....	43
Figure 4–3 Calculating Reconstructed System Spatial Resolution with Scatter	44
Figure 4–4 Volume Sensitivity Cylindrical Phantom	45
Figure 5–1 Source Position for Whole Body Resolution Measurements	49

WORKSHEETS

	Page
Worksheet 4-1 Measurements of FWHM from the Central Point.....	41
Worksheet 4-2 Measurements of FWHM from the Peripheral Points	41

FOREWORD

Reason for Changes

NEMA NU 1-2012 (hereinafter referred to as NEMA NU 1) was developed by the Molecular Imaging Section of the Medical Imaging & Technology Alliance (MITA) of the National Electrical Manufacturers Association (NEMA). Regulations regarding the maintenance of standards by NEMA require that standards be reviewed and, if necessary, updated every five years. Section approval of a standard does not necessarily imply that all section members voted for its approval or participated in its development. At the time of approval, the NEMA NU 1 task force was composed of the following members:

NEMA NU 1 Task Force

- Chuanyong Bai, Digirad Corporation
- Samy Bross, Spectrum Dynamics
- Michael Casey, Siemens Healthcare
- James Hugg, Gamma Medica, Inc.
- Steve Kohlmyer, GE Healthcare
- Jeffrey Kolthammer, Philips Healthcare
- Weidong Luo, Naviscan
- Ron Malmin, Siemens Molecular Imaging
- Navdeep Ojha, Philips Healthcare
- Trudy Rempel, Siemens Healthcare
- Nathaniel Roth, Spectrum Dynamics
- Ling Shao, Philips Healthcare
- Alexander Vija, Siemens Healthcare
- Jason Wiener (Chair), Philips Healthcare
-

NEMA NU 1 Reviewers

- Rafael Puchal Ane, Hospital of Bellvitge, L'Hospitalet de Llobregat, Spain
- Curtis Caldwell, Sunnybrook Odette Cancer Center, Toronto, ONT, Canada
- Anna Celler, University of British Columbia, Vancouver, BC, Canada
- Georges El Fakhri, Massachusetts General Hospital, Boston, MA
- John Engdahl, Bradley University, Peoria, IL
- Tom Lewellen, University of Washington, Seattle, WA
- Stephen Moore, Brigham and Women's Hospital, Boston, MA
- James Scuffham, Royal Surrey County Hospital, Surrey, UK
- Benjamin Tsui, Johns Hopkins Medical Institutions, Baltimore, MD
- Tim Turkington, Duke University, Durham, NC

In the preparation of NEMA NU 1, input of users and other interested parties has been sought and evaluated. Inquiries, comments, and proposed or recommended revisions should be submitted to the relevant MITA product section at the following address:

Vice President
Medical Imaging & Technology Alliance (MITA)
1300 North 17th Street, Suite 900
Rosslyn, Virginia 22209

Changes to the standard

The categorization of tests by section as either “Primary” or “Secondary” tests of system performance remain absent in this revision (consistent with the 2007 version). The relative importance of individual tests to the characterization of Gamma camera systems is left to the manufacturers and user community to determine. The use of “appropriate clinical mode” is also retained, which might affect energy windows and variable count rate operation with respect to the 2001 standard.

As originally scripted, the NEMA NU 1 standard was intended to apply to standard Gamma cameras that utilize large-area-crystal detectors. The 2001 and 2007 versions acknowledged the emergence of discrete pixel detectors, such as pixelated NaI(Tl) and CsI(Tl) detectors, and direct conversion detectors, such as CdTe and CdZnTe (CZT) detectors. Notes regarding the applicability of each test procedure to discrete pixel detectors that were placed near the beginning of each test in the 2007 version are retained in this version.

The NEMA NU 1 task force recognizes, however, that novel SPECT systems are now being deployed that might be characterized by different design constraints or performance capabilities. As a result, many of the tests within the NEMA NU 1 standard define setup or imaging conditions that may not be relevant to describing the delivered quality or performance of new or novel SPECT systems now on the market.

Standard Gamma cameras were developed to deliver large-area planar images with high spatial fidelity to a parallel projection view. The prescribed image evaluation steps for uniformity were chosen to describe this type of design and may not be appropriate for novel SPECT systems that employ advanced reconstruction techniques and algorithms and for which a large-area uniform projection image is not expected.

Similarly, the prescribed image evaluation steps for reconstructed resolution, which dictate filtered back-projection, were chosen as appropriate to the standard Gamma camera design. This technique may not be appropriate for novel SPECT systems that employ advanced reconstruction algorithms and where filtered back-projection is not the preferred or appropriate reconstruction technique.

Standard Gamma cameras deliver a relatively low count rate per unit detector area that limits the total counts and statistical significance in realistic acquisition scenarios. The prescribed image setup and processing steps in NEMA NU 1 were often chosen to minimize the impact of low image count density while preserving relevant performance metrics.

Standard Gamma cameras were designed using PMT optics. The prescribed image setup and processing steps for uniformity were chosen to preserve spatial variations with a scale length of a PMT and may not be appropriate for SPECT systems that employ alternate detection methods.

The NEMA NU1 task force also recognizes that there would be an advantage to reform the standard metrics in order to adequately evaluate the performance of new, non-standard SPECT systems. As of this revision, however, there is no consensus on how to best adapt the NEMA NU 1 standard in order to best characterize novel SPECT system performance.

Changes to definitions and test procedures

Intrinsic Count Rate Performance in Air (Section 2.6) was amended to include the details for an alternate method (Copper Plates Method). This method was included by reference in the 2007 version. The traceability of this method to the Decaying Source Method must be demonstrated by the manufacturer if the Copper Plates Method is to be used.

Intrinsic Count Rate Performance and System Count Rate Performance were amended to accommodate detectors with very high count rate performance (i.e., low dead-time losses, typical of pixelated detectors). Instead of reporting observed count rate at maximum and 20% loss, alternatively, the maximum count rate tested and percentage count loss at that rate can be reported.

Intrinsic Energy Resolution (Section 2.3) was amended to allow an isotope other than Co-57 to determine the keV per channel calibration factor.

Intrinsic Flood Field Uniformity (Section 2.4) was amended for clarity and to include measurements of defective pixels and defective pixel clusters for discrete pixel detectors. The term “defective pixel” was added to the list of defined terms.

Edits for correctness or clarity, of either the test phantom or the source strength, were made to the 2007 version: Sections 2.5 (Multiple Window Spatial Registration), 3.3 (System Planar Sensitivity and Collimator Penetration and Scatter), 4.1 (System Alignment) and 4.2 (SPECT Reconstructed Spatial Resolution without Scatter). In particular, the diameter of the test phantom for System Planar Sensitivity and Collimator Penetration and Scatter is now allowed to be specified by the manufacturer not only for small FOV cameras but also for cameras with focusing collimators.

CAUTION—Persons using this measurement standard must be in compliance with all applicable federal and state regulations (Ref: NRC Regulatory Guide 10.8, *Guide for the Preparation of Applications for Medical Programs*) for the use, handling, and possession of radioactive material.

The purpose of NEMA NU 1 is to provide uniform criteria for the measurement and reporting of Gamma camera performance parameters by which a manufacturer may specify his device and, when doing so, reference “NEMA Standards Publication NU 1-2012, Performance Measurements of Gamma Cameras.” NEMA NU 1 does not establish minimum performance levels.

Specific measurement equipment, as set forth herein, is required in order to accomplish the purpose of this standard: the uniform and accurate specification of performance characteristics. Without this equipment, the measurements performed would be limited, inaccurate, non-quantitative, or too time consuming.

Section 1 SCOPE

This Standards Publication establishes definitions, quantitative measurements of performance characteristics, and reporting techniques for the specification of the following Gamma camera parameters:

- a) Intrinsic Spatial Resolution
- b) Intrinsic Spatial Linearity
- c) Intrinsic Energy Resolution
- d) Intrinsic Flood Field Uniformity
- e) Multiple Window Spatial Registration
- f) Intrinsic Count Rate Performance in Air
- g) Intrinsic Spatial Resolution at 75 kcps
- h) Intrinsic Flood Field Uniformity at 75 kcps
- i) System Spatial Resolution without Scatter
- j) System Spatial Resolution with Scatter
- k) System Planar Sensitivity and Collimator Penetration and Scatter
- l) Detector Shielding
- m) System Count Rate Performance with Scatter
- n) System Alignment
- o) SPECT Reconstructed Spatial Resolution without Scatter
- p) SPECT Reconstructed Spatial Resolution with Scatter
- q) System Volume Sensitivity
- r) Detector-Detector Sensitivity Variation
- s) Whole-body System Spatial Resolution without Scatter

The following types of medical radionuclide imaging instruments are included in this standard:

- a) Single detector, single crystal planar Gamma cameras
- b) Single detector, single crystal tomographic Gamma cameras
- c) Multiple detector planar and tomographic Gamma cameras
- d) Whole-body Gamma camera devices
- e) Discrete pixel detector Gamma cameras

The following types of medical radionuclide imaging instruments are not included in this standard:

- a) Coincidence imaging Gamma cameras or systems (these are covered by NEMA NU 2-2012, "Performance Measurements of Positron Emission Tomographs").
- b) All medical radionuclide imaging devices not included above.

1.1 DEFINITIONS

absolute linearity: The maximum distortion or displacement of the X and Y image location with respect to the actual source location over the Gamma camera field of view (FOV).

appropriate clinical mode: All tests shall be performed in a clinically consistent mode of operation with appropriate energy, linearity and uniformity corrections, pixel size, and photopeak window being employed. The count rate mode employed during tests shall be the same mode as used clinically under the same count rate conditions.