

NEMA MS 4-2006

---

# Acoustic Noise Measurement Procedure for Diagnostic Magnetic Resonance Imaging Devices



**NEMA MS 4**

**ACOUSTIC NOISE  
MEASUREMENT  
PROCEDURE FOR  
DIAGNOSTIC MAGNETIC  
RESONANCE IMAGING  
DEVICES**



**NEMA Standards Publication MS 4-2006**

*Acoustic Noise Measurement Procedure  
For Diagnostic Magnetic Resonance Imaging Devices*

*Published by:*

**National Electrical Manufacturers Association**

1300 North 17th Street, Suite 1752  
Rosslyn, Virginia 22209

[www.nema.org](http://www.nema.org)

© Copyright 2006 by the National Electrical Manufacturers Association. All rights including translation into other languages, reserved under the Universal Copyright Convention, the Berne Convention for the Protection of Literary and Artistic Works, and the International and Pan American Copyright Conventions.

## **NOTICE AND DISCLAIMER**

The information in this publication was considered technically sound by the consensus of persons engaged in the development and approval of the document at the time it was developed. Consensus does not necessarily mean that there is unanimous agreement among every person participating in the development of this document.

NEMA standards and guideline publications, of which the document contained herein is one, are developed through a voluntary consensus standards development process. This process brings together volunteers and/or seeks out the views of persons who have an interest in the topic covered by this publication. While NEMA administers the process and establishes rules to promote fairness in the development of consensus, it does not write the document and it does not independently test, evaluate, or verify the accuracy or completeness of any information or the soundness of any judgments contained in its standards and guideline publications.

NEMA disclaims liability for any personal injury, property, or other damages of any nature whatsoever, whether special, indirect, consequential, or compensatory, directly or indirectly resulting from the publication, use of, application, or reliance on this document. NEMA disclaims and makes no guaranty or warranty, express or implied, as to the accuracy or completeness of any information published herein, and disclaims and makes no warranty that the information in this document will fulfill any of your particular purposes or needs. NEMA does not undertake to guarantee the performance of any individual manufacturer or seller's products or services by virtue of this standard or guide.

In publishing and making this document available, NEMA is not undertaking to render professional or other services for or on behalf of any person or entity, nor is NEMA undertaking to perform any duty owed by any person or entity to someone else. Anyone using this document should rely on his or her own independent judgment or, as appropriate, seek the advice of a competent professional in determining the exercise of reasonable care in any given circumstances. Information and other standards on the topic covered by this publication may be available from other sources, which the user may wish to consult for additional views or information not covered by this publication.

NEMA has no power, nor does it undertake to police or enforce compliance with the contents of this document. NEMA does not certify, test, or inspect products, designs, or installations for safety or health purposes. Any certification or other statement of compliance with any health or safety-related information in this document shall not be attributable to NEMA and is solely the responsibility of the certifier or maker of the statement.

## CONTENTS

	Page
Preamble .....	ii
Foreword .....	iii
Introduction .....	iv
Rationale .....	iv
Section 1 REFERENCED STANDARDS AND DEFINITIONS .....	1
1.1 SCOPE.....	1
1.2 REFERENCED STANDARDS .....	1
1.3 DEFINITIONS .....	1
Section 2 DATA ACQUISITION PARAMETERS .....	3
2.1 MAXIMUM GRADIENT ACOUSTIC NOISE (MGAN) METHOD .....	3
2.2 MAXIMUM CLINICAL ACOUSTIC NOISE (MCAN) METHOD.....	3
Section 3 METHODS OF MEASUREMENT .....	5
3.1 TEST HARDWARE .....	5
3.2 HARDWARE SETUP .....	5
3.3 MGAN MEASUREMENT .....	5
Section 4 RESULTS.....	10
4.1 REPORTING RESULTS .....	10
Annex A ACOUSTIC NOISE EXPOSURE STANDARDS .....	11
A.1 MOTIVATION FOR THE REVISED STANDARD .....	11
A.2 COMMENTS ON THE MGAN METHOD .....	11
A.2.1 Rationale for $f_{\max}$ and $3 \times f_{\max}$ .....	11
A.2.2 Measurement of the noise spectrum .....	12
A.2.3 Finding the loudest spot.....	12
A.3 COMMENTS ON THE MCAN METHOD .....	12

## **Preamble**

This is one of a series of test standards developed by the medical diagnostic industry for the measurement of parameters governing the safety of Magnetic Resonance (MR) Imaging systems. These test standards are intended for the use of equipment manufacturers, testing houses, prospective purchasers, and users alike.

Manufacturers are permitted to use these standards for the determination of system performance specifications. This standardization of performance specifications is of benefit to the prospective equipment purchaser. The parameters supplied with each NEMA measurement serve as a guide to those factors which can influence the measurement. These standards can also serve as reference procedures for acceptance testing and periodic quality assurance.

It must be recognized, however, that not all test standards lend themselves to measurement at the installation site. Some test standards require instrumentation better suited to factory measurements, while others require the facilities of an instrumentation laboratory to assure stable test conditions necessary for reliable measurements.

The NEMA test procedures shall be carried out using the normal clinical operating mode of the system. For example, standard calibration procedures and standard reconstruction processes shall be used. No modifications to alter test results shall be used unless otherwise specified in these standards.

The NEMA Technical Committee of the MR Section has identified a set of key magnetic resonance safety parameters. This standard describes the measurement of one of these parameters.

## **Equivalence**

It is intended and expected that manufacturers or others who claim compliance with these NEMA standard test procedures for the determination of safety parameters shall have carried out the tests in accordance with the procedures specified in the published standards.

In those cases where it is impossible or impractical to follow the literal prescription of a NEMA test procedure, a complete description of any deviation from the published procedure must be included with any measurement claimed to be equivalent to the NEMA standard. The validity or equivalence of the modified procedure will be determined by each reader.

## **Uncertainty of the Measurements**

The measurement uncertainty of the safety parameter determined using this standard is to be reported, together with the value of the parameter. Justification for the claimed uncertainty limits shall also be provided by a listing and discussion of sources and magnitudes of error.

## Foreword

This standards publication is classified as a NEMA standard unless otherwise noted. It describes the test conditions and parameters that approximate the worst case acoustic noise levels that a particular magnet/gradient system combination produces when using pulsed gradient waveforms. This standard also describes how the acoustic noise levels are to be measured. In the absence of specific guidelines for sound level exposure with MR imaging equipment, this procedure references the OSHA guidelines for acoustic noise exposure and the IEC standards for sound level meters.

This standards publication has been developed by the Magnetic Resonance Section of the National Electrical Manufacturers Association. User needs have been considered throughout the development of this publication. Proposed or recommended revisions should be submitted to:

Vice President, Engineering Department  
National Electrical Manufacturers Association  
1300 North 17th Street  
Rosslyn, VA 22209

Section approval of the standard does not necessarily imply that all section members voted for its approval or participated in its development. At the time it was approved, the section was composed of the following members:

Computer Imaging Reference Systems—Norfolk, VA  
Fonar Corporation—Melville, NY  
GE Healthcare, Inc.—Milwaukee, WI  
Hitachi Medical Systems America, Inc.—Twinsburg, OH  
Invivo Diagnostic Imaging—Gainesville, FL  
Philips Medical Systems North America—Bothell, WA  
Siemens Medical Solutions, Inc.—Malvern, PA  
Toshiba America Medical Systems—Tustin, CA

## Introduction

Current passing through a wire placed in a magnetic field will generate a force orthogonal to both the direction of the field and the current (Lorentz force). Acoustic noise in MR imaging is caused by motion of the gradient coils and attached structures when a time-varying current is passed through the coils.

At a given magnetic field strength, the loudest noise will occur when pulsed waveforms are applied simultaneously to all three gradient coils at maximum amplitude and at a system-dependent resonant frequency. Though this condition may not be realized for most clinical imaging sequences, it nonetheless represents the worst case condition and is defined here as the Maximum Gradient Acoustic Noise (MGAN) of the system. The Maximum Clinical Acoustic Noise (MCAN) of the system is defined as the clinical imaging condition that produces the greatest acoustic noise. Typically, this occurs with pulse sequences that use high gradient current duty cycles and high gradient current amplitudes near system-dependent resonant frequencies.

To determine the acoustic noise of a system either the MGAN or the MCAN can be measured and reported. The choice of using the MGAN or the MCAN is left to the discretion of the user.

## Rationale

The sound generated by an MR system usually consists of a series of repetitive impulses. The relevant safety parameters required to characterize such a noise are the unweighted peak sound pressure level ( $L_{\text{peak}}$ ) and the time integral of the A-weighted sound pressure level ( $L_{\text{Aeq}}$ ). In MR applications the peak sound pressure level is dependent on the peak amplitude of the individual pulses while the time integral of the A-weighted sound pressure level is dependent on the continuous exposure to a series of such pulses. This document describes methods to measure the worst case  $L_{\text{peak}}$  and  $L_{\text{Aeq}}$  of an MR System.

## Section 1 REFERENCED STANDARDS AND DEFINITIONS

### 1.1 SCOPE

The purpose of this NEMA Standards Publication is to provide methods to determine the acoustic noise level of an MR system. Two measurement procedures are defined, Maximum Gradient Acoustic Noise (MGAN) and Maximum Clinical Acoustic Noise (MCAN).

This procedure has been designed for measuring peak sound pressure levels (SPL) up to 140 dB. Above 140 dB, the use of more sophisticated equipment and methods may be required.

### 1.2 REFERENCED STANDARDS

The following publications are adopted as indicated by reference in this standards publication. The mailing address of each referenced organization is also provided.

#### International Electrotechnical Commission

1, Rue de Varembe  
Geneva, Switzerland

IEC 61672-1, 2002

*Sound Level Meters*

IEC 60601-2-33, 2002

*Particular requirements for the safety of magnetic resonance equipment for medical diagnosis*

#### Occupational Safety and Health Administration

Department of Labor  
200 Constitution Avenue, N.W.  
Washington, D.C. 20210

29CFR-1988, Part 1910.95

*Occupational Noise Exposure*

#### American National Standards Institute

1430 Broadway  
New York, NY 10018

ANSI S1.4-1983

*Sound Level Meters*

### 1.3 DEFINITIONS

**SPL (Sound Pressure Level):** The SPL is defined as ten times the common logarithm of the ratio of the square of the measured sound pressure to the square of the standard reference pressure of 20 micropascals.

**A-weighting:** This refers to SPL frequency weighting. The ear does not respond uniformly to all frequencies. SPL measurements made with an A-weighting correspond to noise levels that are similar to those heard by the human ear (IEC 61762).